

Corrigendum

Sub : Limited Tender for Supply, Delivery, Installation & Commissioning of 2(Two) No. Mechanical Ventilator , under buy back offer of 1 (one) No for use in ICU at Centenary Hospital of Kolkata Port Trust. Tender No. MED / PR / LMT / 289/ 18-19/ 819 Dated:06/7/2018 & E- Tender No: KoPT/Kolkata DockSystem/MED/7/18-19/ET/208

- A. **CLOSING DATE** of e- tender will be 10/12/2018 at 15 hrs in place of existing entry 29/11/2018 at 15 hrs
 - B. **OPENING DATE** of e- tender will be 10/12/2018 at 15:30 hrs in place of existing entry 29/11/2018 at 15:30 hrs.
 - C. **Technical Specification Appendix B (Page 23)** of the referred tender has been revised as follows " in place of the existing entry.
- 1. BATTERY BACKUP FOR THE ENTIRE VENTILATOR SYSTEM FOR AT LEAST ONE HOUR(ORIGINAL SPECIFICATION WAS BACKUP FOR THREE HOURS) OR MORE.(ALSO"INBUILT/INTERNAL" TERM IS OMITTED).
 - 2. THE VENTILATOR CAN BE USED FOR INTRA HOSPITAL USE ONLY(ORIGINAL SPEC .WAS FOR BOTH INTRA AND OUT SIDE HOSPITAL(TRANSPORT) PURPOSE.
 - 3. IN VENTILATOR MODES " OTHER OPTIONAL MODES" CLAUSE IS OMITTED,ONLY MENTIONED MODES WILL BE AVAILABLE.
 - 4. TIDAL VOLUME RANGE IS SET AT 50-2000 ML OR BETTER(ORIGANAL SPEC. WAS 20-2000 ML).
 - 5. FLOW TRIGGER SENSITIVITY SET AT 1.0-10 ML/MIN. OR BETTER(ORIGINAL SPEC WAS 0.1 ML/MIN- 20ML/MIN)
 - 6. REAL TIME DISPLAY OF MINIMUM TWO CURVES ON SCREEN (ORIGINAL SPEC. WAS MINIMUM THREE CURVES IN REAL TIME).
 - 7. USER INTERFACE CONNECTOR COMPATATIBILITY REQUIRED. (ORIGINAL SPEC .WAS USB AND R 145).
 - 8. INSPIRATORY SLOPE 20-200 CM H₂O(ORIGINAL SPEC. WAS 10-200 CM H₂O).
 - 9. PEAK FLOW SETTINGS -0-180L/MIN OR BETTER (ORIGINAL SPEC. WAS 0-240 ML/MIN OR BETTER)

10. EXP.FLOW SENSOR AND EXP. VALVE CASSETTE SHOULD BE STEAM/AUTOCLAVE STERILIZABLE.(ORIGINAL WAS FOR BOTH DISPOSABLE AND STEAM AND AUTOCLAVE STERILIZABLE).

ACCESSORIES-

1. PATIENT BREATHING CIRCUIT(NON DISPOSABLE/REUSABLE) -TWO(02) NOS OF ADULT AND ONE (01) No PEDIATRIC ,WITH EACH VENTILATOR.
2. 2.REUSABLE NIV MASK -LARGE SIZE 01,MEDIUM SIZE 01 AND SMALL 01 NO WITH EACH VENTILATOR.
3. AUTO HEATED HUMIDIFIER(ORIGINAL MAKE) ONE (01) EACH WITH CIRCUIT ONE EACH WITH VENTILATOR.
4. HEPA FILTER FOR TURBINE -05(FIVE) FOR EACH VENTILATOR.
5. FLOW SENSOR/EXP.CASSTETT -05(FIVE) FOR EACH VENTILATOR.

END OF CORRIGENDUM.

Date: 26/11/2018