

कोलकाता पत्तन न्यास

KOLKATA PORT TRUST

OFFICE OF THE CHIEF MEDICAL OFFICER

CENTENARY HOSPITAL, MEDICAL DEPARTMENT

1, DIAMOND HARBOUR ROAD, KOLKATA - 700053

TELE FAX NO: (033)-2401 4503, PHONE NO: (033)-2401 4094 (Extn-841)

Advertised Tender for Supply, Delivery, installation and commissioning of Assorted Pathology Equipments at Centenary Hospital, KolkataPort Trust

Tender No : MED / PR / ADVT / 264 / 16-17/ 1835 Dated 02/02/2017

CORRIGENDUM

Technical Specification Annexure X (Page 14 of 18) of Tender No MED/ PR/ ADVT/264/ 16-17/ 1835 dated 30/1/20167 (KOPT/Kolkata Dock System/ MED /12/ 16-17 / ET / 156) for Laboratory Centrifuge Machine following entries are added after serial no 18

A. <u>Laboratory Centrifuge Machine</u>

- 19. Required RPM should be in the range of 1000 RPM 5000 RPM
- 20. Sample Volume 3ml -5 ml
- 21. No of tubes required 24 nos.

This Corrigendum should be treated as part of the Tender.

(Dr. Sumita Roy.) Sr. Medical Officer- Pathology

Date: 16/2/2017 For Chief Medical Officer