



# कोलकाता पत्तन न्यास

चिकित्सा विभाग,सेन्टेनरी अस्पताल

1, डायमंड हार्बर रोड, कोलकाता-700 053

## KOLKATA PORT TRUST

MEDICAL DEPARTMENT, CENTENARY HOSPITAL

1, DIAMOND HARBOUR ROAD, KOLKATA-700053

[ISO 9001: 2008 Certified]

[TELE FAX NO : \( 033\)-2401 4503](tel:03324014503), [PHONE NO: \(033\)- 2401](tel:03324013817)

[3817 / 4094 \(Extn-841\)](tel:03324013817)

e-mail : [cmo@kolkataporttrust.gov.in](mailto:cmo@kolkataporttrust.gov.in)



### CORRIGENDUM-I

Sub:-Notice Inviting Open Tender for Supply, Delivery, Installation & Commissioning of 2(two )Nos. Manual External Biphasic Defibrillator for use in Operation Theatre of Centenary Hospital of KoPT..Tender No. **MED / PR / ADVT / 309R/ 19-20 / 1468 Dated: 22/11/2019**  
Tender ID 2019\_KoPT\_521644 of CPPP portal.

A. In response to pre-bid meeting held on 10/12/2019 Queries has been answered as follows :-

SL	Existing Tech Specification	Vendor's request	Reply
General	BATTERY- The unit should have a built in rechargeable battery with at least 3 hours of monitoring times or 100 discharges of at least 200 J with fully charged batteries. A low battery indicator should be displayed on the screen	The unit should have a built in rechargeable battery with at least 4 hours of monitoring times or 100 discharges of at least 300 J with fully charged batteries. A low battery indicator should be displayed on the screen	No Change
do	The unit should have a charge time of less than 7 sec to 200 J (for at least first 10 charges to 200 J).	The unit should have a charge time of less than 8 sec to 300 J (for at least first 10 charges to 200 J).	No Change
PRINTER -	The printer should display time, date, heart rate, heart rate limits, event marker, ECG parameters, defibrillation mode, selected and delivered energy with patient impedance and Hospital name (preferably).	<b>Print annotations</b> :The printer should display time, date, heart rate, heart rate limits, event marker, ECG parameters, defibrillation mode, selected and delivered energy with patient impedance and Hospital name (preferably). <b>selected and delivered energy with patient impedance and Hospital name (preferably). Deleted this high lighted point</b>	No Change
STANDARD ACCESSORIES	Internal Paddles- 1	To be deleted.	Deleted
		<b>To add :1) screen size 5 inch. 2.) output energy should be 5-360 J.</b>	No Change



# कोलकाता पत्तन न्यास

चिकित्सा विभाग,सेन्टेनरी अस्पताल

1,डायमंड हार्बर रोड,कोलकाता-700 053

## KOLKATA PORT TRUST

MEDICAL DEPARTMENT,CENTENARY HOSPITAL

1, DIAMOND HARBOUR ROAD,KOLKATA-700053

[ISO 9001: 2008 Certified]

[TELE FAX NO : \(033\)-2401 4503](tel:(033)-24014503), [PHONE NO: \(033\)- 2401 3817 / 4094 \(Extn-841\)](tel:(033)-24013817)

[e-mail : cmo@kolkataporttrust.gov.in](mailto:cmo@kolkataporttrust.gov.in)



--	--	--	--

B. Tender schedule has been revised as follows :-

	tender schedule	Existing	Amended
10	Date of closing of Online e-tender for submission of Techno-Commercial Bid & Price Bid at CPPP Portal	19 / 12/ 2019 at 15.00 hrs.	26 / 12/ 2019 at 15.00 hrs.
11	Date & time of opening of (Techno-Commercial Bid + Price bid) at CPPP Portal	20/ 12/ 2019 at 15.30 hrs.	27/ 12/ 2019 at 15.30 hrs.

16-12-2019